



REPUBLIC
OF
KENYA



Poverty Reduction Through Sustainable NRM



Investing in rural people

MINISTRY OF WATER, SANITATION & IRRIGATION

UPPER TANA CATCHMENT NATURAL RESOURCES MANAGEMENT PROJECT (UTaNRMP)

PO Box 996-60100 EMBU Tel: 068-2231376 E-mail: utanrmp@gmail.com info@utanrmp.or.ke www.utanrmp.or.ke

Addendum no 2

CLARIFICATION OF TENDER No: MWSI/UTaNRMP/NCB/001/2022-23

Upper Tana Catchment Natural Resources Management Project (UTaNRMP) would like to make the following clarifications for Tender Ref No: MWSI/UTaNRMP/NCB/001/2022-23

	Issue	Response/Clarification
1	Refrigerator 300ltr., a refrigerator cannot have a temp range from 5°C-99,9°C pls. clarify.	Refrigerator temperature is 2 to 5 (2-5 ⁰ C) degrees Centigrade in the main compartment. The freezer compartment is at minus 18 degrees Centigrade
2	Incubator min 60 litre: this size does not exist. I suggest 50 Ltr. or 105 ltr. Also pls clarify if natural convection or forced convection is required	The incubator should be 50 Ltr minimum. Must be Natural convection.
3	Thermometer, the description is not clear can it be that you need a testo 174 T - Mini-Temperatur-Datenlogger?	Thermometer should have a minimum of minus 40 degrees Centigrade as the lower temperature and plus 70 degree higher limit.
4	Pos 5 Thermohygrometer, please provide detailed	Thermo - hygrometer to measure and display temperature and humidity. For the temperature - range is between zero to +70 degrees Centigrade, accuracy of plus or minus 1degrees Centigrade and resolution of 0.1 degrees Centigrade. For the Humidity range is between 2 to 98% RH, Accuracy of plus or minus 5% RH and resolution of 1% RH.
5	Pos 6 specs fume hood are missing the application to define the carbon filter Please advise	Non ducted – must have a carbon filter which are activated carbon filters
6	Pos 8 Gel doc, please provide more details about size ect.	It should include high resolution camera minimum 9 megapixels, minimum 12.1 inch touch screen, and

		should include software to visualize and capture images following Gel electrophoresis
7	Lot 2 pos 6, what do you mean by baby jar?	These are 250 ml glass bottles or jars with autoclavable plastic lids to use as TC Culture vessels
8	Lot 3, please be so kind and provide Cas number or item code of a manufacturer such as Merck or sigma Adrich. the reason is that there are different grades which are different in price.	The specifications are general without specifying manufacturer to avoid bias

Bidders are informed that this addendum forms part of the tender documents. All other terms and conditions remain the same.

**PROJECT COORDINATOR
UPPER TANA CATCHMENT NATURAL RESOURCES MANAGEMENT PROJECT.**